

Specification Sheet

Specification Number: 8015VB-170S
Revision: 03
Date Issued: 18-Nov-22

1. Part Number: 8015VB-170S

2. Photograph:



3. Product Information:

Description: SteriWare® Very Narrow Micro Spatula 0.3ml Blue - Sterile

Use: Single use, powder spatula

Material of Construction: Polypropylene – virgin, blue

Conforms to FDA 21 CFR 177.1520
Conforms to EU Regulations 10/2011
Conforms to EC Regulations 1935/2004

• BSE/TSE free

Method of Construction: Single piece, injection moulded

Overall Length 170mm

Blade Length: 40mm

Blade Width: 4mm

Nominal Weight: 4g

Nominal Volume: 0.3ml

Moulding Environment: Class 100,000 Cleanroom
Primary Packing Environment: Class 10,000 Cleanroom

Individually Bagged? Yes (heat sealed PE/PET bag)
Method of Sterilisation: Gamma Irradiation (25 to 45 kGy)

Number of Samplers per Box: 100

Recommended Storage Conditions: Dry and ambient temperature

Shelf Life: 5 years from the date of manufacture
ATEX: This device has not been ATEX tested



4-5 Forge Mills Park, Coleshill. Warwickshire. B46 1JH. UK. Tel: +44(0)1675 466992 Email: info@sampling.com Web: www.sampling.com



Specification Sheet

Specification Number: 8015VB-170S Revision: 03 Date Issued: 18-Nov-22

4. Change Control: Sampling Systems operates a policy of continual improvement.

To be kept informed of all major changes to any product in the

SteriWare range please email your contact details to

change.control@sampling.com

5. Country of Origin: Made and packed in the UK.

6. Repeated Use: This is a single-use, disposable sampler. Repeated use will lead to

failure.

7. Approval: Signed:

Name: Kuldeep Flora

Position: Quality Manager

Date: 18-Nov-22